



### Indicator Dot Tags

**INTENDED USE:**

Indicator Dot Tags are used to identify instruments, trays, carts and other equipment as they move from the sterilization to staging areas of surgery. Indicator dots on tags change color to indicate exposure to steam sterilization.

**CAUTION:**

Store at 20-25°C (68-77°F), excursions 15-30°C (59-86°F) allowed. Relative humidity range is 30-70%. Significant changes in storage conditions for prolonged periods can have an adverse effect on the product. Extreme storage conditions such as exposure to direct sunlight and/or storage on top of or near a heat source should be avoided. Store away from hydrogen peroxide and other sterilant.

**INSTRUCTIONS FOR USE:**

1. Tags do not need to be cleaned or disinfected before use.
2. Tags can be sterilized according to the parameters detailed below. Tags are single-use and can be sterilized only once.
3. Verify that the indicator dot changes from blue to dark. Typical color change will result in dark brown to black after exposure to steam sterilization.
4. Tags can be broken and removed by hand or with the use of cutters.

**STERILIZATION:**

The Tags are provided non-sterile. They have been validated for steam sterilization at the following cycle parameters:

<i>Cycle Type</i>	<i>Temperature</i>	<i>Minimum Time – Full Cycle</i>
Pre-Vacuum	132° C (270° F)	4 minutes

**DISPOSAL:**

The tags should be disposed of after use; there are no special disposal instructions.

 **MANUFACTURER: KEY SURGICAL LLC**  
 8101 WALLACE ROAD / EDEN PRAIRIE, MN 55344 / USA  
 PH: 800.541.7995 KEYSURGICAL.COM

**REF** CATALOG NUMBER

**LOT** LOT NUMBER  **CONSULT INSTRUCTIONS FOR USE**

 **USE-BY DATE**  **DO NOT RE-USE**  **CAUTION**