


**Instructions for Use:** Instrument Repair tags

<b>Brand Name of Product</b>	Instrument Repair Tags
<b>Generic Name of Product</b>	Instrument Repair Tags
<b>Product Code Number(s)</b>	994-G, 994-R, 994-S, 5030 WT CLN
<b>Purpose of Product</b>	To identify instruments requiring repair, sharpening or “good” with brightly colored coded tags.
<b>Range of Applications for Product</b>	Surgical instruments and other medical devices requiring staff attention
<b>Key Specifications of Product</b>	<ul style="list-style-type: none"> <li>• Sealable and breakable by hand.</li> <li>• Tamper evident – once broken, cannot be resealed.</li> <li>• Color-coded with related printed message. <ul style="list-style-type: none"> <li>○ Green – Good</li> <li>○ Red – Repair</li> <li>○ Yellow – Sharpen</li> <li>○ White Pulltight – “Clean”</li> </ul> </li> <li>• Manufactured of lightweight and durable plastic.</li> </ul>

**Shipping & Storage**

<b>Shipping Conditions &amp; Requirements</b>	
<b>Storage Conditions</b>	
<b>Packaging Contents</b>	100 tags per box
<b>Shelf Life</b>	

**Instructions for Using Product**

<b>Description of Use (s)</b>	Once tags are applied, they must be broken to be removed. Replace difficult to see pins and suture.
<b>Preparation for Use</b>	
<b>Diagrams (drawings, pictures)</b>	
<b>Steps for Use of Product</b>	<b>Apply &amp; Remove</b> <ol style="list-style-type: none"> <li>1. Place arrow through slot of item to be sealed.</li> <li>2. Gently push arrow into locking barrel until it engages locking teeth.</li> <li>3. Gently push on padlock seal to ensure it has been properly locked.</li> <li>4. To remove, hook forefinger under locking barrel and pull laterally away from the slot. <b>DO NOT PULL DOWNWARD</b></li> </ol>
<b>Interpretation of Results</b>	
<b>Contraindications of Test Results</b>	
<b>Documentation</b>	
<b>Special Warnings and Cautions</b>	<b>DO NOT PULL DOWNWARD</b>
<b>Disposal</b>	

**Reprocessing Instructions**

<b>Point of Use</b>	
<b>Preparation for Decontamination</b>	
<b>Disassembly Instructions</b>	
<b>Cleaning – Manual</b>	
<b>Cleaning – Automated</b>	
<b>Disinfection</b>	
<b>Drying</b>	
<b>Maintenance, Inspection, and Testing</b>	

<b>Reassembly Instructions</b>	
<b>Packaging</b>	
<b>Sterilization</b>	Tags are materially compatible with steam, EtO and gas-plasma sterilization.
<b>Storage</b>	
<b>Additional Information</b>	

<b>Related Healthmark Products</b>	Scissor Tests
<b>Other Product Support Documents</b>	ProSys™ Brochure, ProSys™ Price List
<b>Reference Documents</b>	
<b>Customer Service Contact</b>	Healthmark Industries Company, Inc 18600 Malyn Blvd. Fraser, MI 48026 1-586-774-7600 <a href="mailto:healthmark@hmark.com">healthmark@hmark.com</a> <a href="http://hmark.com">hmark.com</a>

2021-03-08 msmith

DISTRIBUTED BY: KEY SURGICAL  
8101 WALLACE ROAD / EDEN PRAIRIE, MN 55344 / USA  
(US) PH: 800.541.7995 KEYSURGICAL.COM  
(UK) TEL: +44 (0)1628 810626 KEYSURGICAL.CO.UK  
(EU) TEL: +49 4363 905900 KEYSURGICAL.DE