

INSTRUCTIONS FOR USE

TRANSPORTATION COVER BAGS

Name: Transportation Cover Bags and Transportation Cover Dry Bags

Product Code: TCOVERxxx, TCOVERDRYxxx

Name of Manufacturer: Clinipak Ltd
Address: Beech House

aress. Decem riouse

Knaves Beech Business Centre

Loudwater
High Wycombe

Bucks HP10 9SD

Tel: +44 (0)1628 810626 Fax: +44 (0)1628 810760

Website: clinipak.co.uk

Intended Use

Clinipak Cover Bags and Transportation Cover Dry Bags are single use, nonwoven, autoclavable bags which are designed to be used as outer transportation wraps in sterile barrier systems.

Choice Dry Bag Blue and Choice Dry Bag Green Bags are intended to be used by professional healthcare personnel working in sterile services (or similar) departments.

Single Use

Clinipak Cover Bags and Transportation Cover Dry Bags are single use only.

Sterilisation Compatibility

Clinipak Cover Bags are compatible with steam, ETO and H2O2/Plasma sterilization processes Transportation Cover Dry Bags are compatible only with steam and ETO sterilization processes.

Storage Conditions

It is recommended that Clinipak Cover Bags and Transportation Cover Dry Bags are stored at 10-35°C, 20-80%RH. If this is not possible, it should be prevented from getting wet or from being exposed to high temperatures (>40degC).

Precautions

Clinipak Cover Bags and Transportation Cover Dry Bags should be stored away from any sources of ignition. Transportation Cover Dry Bags should not be used in H202/Plasma sterilization processes

Disposal

Unused Clinipak Cover Bags and Transportation Cover Dry Bags can be disposed of in the plastic recycling stream.

Used Clinipak Cover Bags and Transportation Cover Dry Bags should be disposed of by taking into account possible microbial contamination. The user should reference their own procedures in this matter.

Document Ref: IFU-3028
Page 1 of 3

Effective Date: 13 Feb 2018

TMP-IFU-V01: Effective 13 Feb 2018

Version No: 01

CLINIPAK LTD



How to Use Clinipak Cover Bags and Transportation Cover Dry Bags

- 1) Wrap the basket or tray containing the medical devices to be sterilized in 1 or more sterilization barrier wraps:
- 2) Place the Wrapped basket or tray inside the Bag:



3) Fold over the lip of the bag:



4) Fold-over each corner of the Lip of the Bag:



Document Ref: IFU-3028

Page 2 of 3

Effective Date: 13 Feb 2018

TMP-IFU-V01: Effective 13 Feb 2018

Version No: 01

CLINIPAK LTD



5) Fold-over the Lip 3 times:



6) Place Autoclave tape over end:



Document Ref: IFU-3028

Page 3 of 3

Effective Date: 13 Feb 2018

TMP-IFU-V01: Effective 13 Feb 2018

Version No: 01