

Identification Tape

INTENDED USE: Key Surgical® Identification Tape is intended to adhere to a stainless steel surgical instrument to help with identification, organization, management, proper processing, and packaging of such using various colors, patterns and/or alphanumeric coding.

PRECAUTION:

Identification tape is not intended for implantation.

INSTRUCTIONS FOR USE:

1. Clean fingers with isopropyl alcohol to remove oils or wear gloves.
2. Clean and dry application area of instrument to remove any residue that may exist.
3. Wipe area with alcohol; allow alcohol to dry.
4. Firmly apply tape; do not stretch. Tape should be applied to the nonworking end of the instrument. Application should not impede instrument functionality.
5. Wrap tape 1.5 times on a stainless steel instrument. Tape should lay flat without gaps.
6. Steam sterilize the instrument to bond the adhesive.

INSPECTION:


Inspect tape prior to each use. Identification tape is not intended as a permanent mark and will discolor, break, chip, or flake over time. Replace as soon as these are noticed.

STERILIZATION: Identification tape is provided non-sterile and has been validated for steam sterilization efficacy including sterilization of the underlying instrument surface using the following full cycle parameters:

Cycle Type	Temperature	Minimum Time – Full Cycle
Gravity	121° C (250° F)	30 minutes
Pre-Vacuum	132° C (270° F)	4 minutes
Pre-Vacuum	134° C (273° F)	3 minutes
Immediate Use	132° C (270° F)	3 minutes

DISPOSAL:

There are no special disposal instructions.

 **MANUFACTURER: KEY SURGICAL LLC**
 8101 WALLACE ROAD / EDEN PRAIRIE, MN 55344 / USA
 PH: 800.541.7995 KEYSURGICAL.COM

REF

CATALOG
NUMBER

LOT

LOT
NUMBER



DO NOT RE-USE



USE-BY DATE